



29th Annual Scientific Sessions

May 14-17, 2008 • San Francisco, CA USA

[Print this Page for Your Records](#)[Close Window](#)**Control/Tracking Number:** 08-AB-5479-HRS**Activity:** Abstract Submission**Current Date/Time:** 1/4/2008 9:59:16 AM**Evoked Response Characteristics from Different Defibrillation Leads for Capture Verification in ICD/CRT-D**

Author Block: Steve Greer, MD, Jonathan Lowy, MD, Mohammad Saeed, MD, John L. Schutzman, MD, Raveen Bazaz, MD, Anne Swearingen, RN, MSN, MBA, CCRP, Yanting Dong, PhD, Kenneth N. Hayes, BS and Michael R. Gold, MD, PhD. Arkansas Cardiology, Little Rock, AR, St. Joseph Hospital, Bellingham, WA, St. Lukes Episcopal Hospital/ Texas Heart Institute, Houston, TX, The Care Group, LLC, Indianapolis, IN, University of Pittsburgh Medical Center, Pittsburgh, PA, Boston Scientific, St. Paul, MN, Medical University of South Carolina, Charleston, SC

Abstract: Introduction: Automatic pacing Threshold (AT) testing with output adjustment may simplify device follow-up and prolong device longevity. An RV AT algorithm based on RVcoil to Can evoked response (ER) sensing was developed for ICD/CRT-D. This analysis evaluated whether ER characteristics were related to defibrillation lead type. Methods: Patients (pts) scheduled for ICD/CRT-D implant, replacement or upgrade were enrolled. An external pacing system (Boston Scientific, St. Paul, MN) with customized software was used to perform threshold tests and data acquisition. RV manual threshold and AT tests were conducted. The RVcoil to Can ER was evaluated offline using minimum ER amplitude (ERmin), maximum pacing artifact (ARTmax), signal to artifact ratio (SAR = ERmin/ARTmax), ER peak timing (tER) stability (delta T = tERmax - tERmin). Signal quality sufficient for capture detection was defined as SAR > 2 and delta T < 45ms. Results: Data from 106 pts (80 M/26 F, 67±11 years, LVEF: 28% ± 11%, 64 ICD/42 CRT-D) were analyzed. Chronic leads had smaller ER amplitude and artifact (p < 0.05) than those of acute leads. No statistical difference was seen in any ER measures between dedicated and integrated configuration. SAR > 2 was observed in all pts. However, 3 pts had delta T ≥ 45ms, 2 with acute and 1 with dedicated lead. Conclusions: Both acute/chronic, dedicated/integrated defibrillation leads studied had similar and acceptable RVcoil to Can ER signals for capture verification. The RV AT algorithm using RVcoil to Can ER sensing may function for a wide range of defibrillation leads.

| Lead Type | # of Pts with acceptable ER | ERmin (mV) | ARTmax (mV) | SAR | Delta T (ms) |
|------------|-----------------------------|------------|-------------|-----------|--------------|
| Acute | 78/80 | 4.13±1.64 | 0.22±0.14 | 26.7±21.1 | 16.5±9.67 |
| Chronic | 25/26 | 3.32±1.26 | 0.13±0.08 | 34.2±24.7 | 16.0±8.86 |
| Dedicated | 29/30 | 3.89±1.49 | 0.20±0.14 | 31.1±27.1 | 20.3±8.59 |
| Integrated | 74/76 | 3.95±1.63 | 0.20±0.13 | 27.6±20.0 | 14.8±9.37 |

:
Author Disclosure Information: S. Greer, Medtronic, M,I; Boston Scientific, M,I; St. Jude Medical, M,I; J. Lowy, None; M. Saeed, None; J.L. Schutzman, None; R. Bazaz, None; A. Swearingen, Boston Scientific, S,K; Y. Dong, Boston Scientific, S,D; Boston Scientific, S,K; K.N. Hayes, Boston Scientific, S,K; M.R. Gold, Boston Scientific, S,A; Medtronic, S,A; Boston Scientific, S,I; Medtronic, S,I; Boston Scientific, S,J; Medtronic, S,J.

Category Selection (Complete): 24 Bradycardia: Device Technology

Keyword (Complete): +Leads, defibrillation

Additional Abstract Information (Complete):

Presentation Preference: Oral or Poster

I am interested in submitting an abstract for one of the Late-Breaking Trials sessions: No

Abstract Awards (Complete):

None : True

Status: Complete

If you have any **technical questions** or experience any problems with the online submission site, please contact Technical Support at support@abstractsonline.com or call 217-398-1792. Technical Support hours of operation: M-F 9am-5pm CST.

If you have any questions concerning the 2008 Scientific Sessions, Feel Free to contact Tim Gregory, 202-464-3438 or scientific.sessions@hrsonline.org.

Powered by [OASIS](#), The Online Abstract Submission and Invitation System SM
© 1996 - 2008 [Coe-Truman Technologies, Inc.](#) All rights reserved.